

Orthotist's Name:

Ankle Foot Orthosis (AFO)

The contact details for our department are:

*The Orthotics Department
SMART Building
Astley Ainslie Hospital
133 Grange Loan Edinburgh
EH9 2HL
Tel: 0131 537 9418*

*Orthotics Office Ground Floor
St. John's Hospital
Howden Road West
Livingston
EH546PP
Tel: 01506 522108*

*Orthotics Office
1st Floor Old Out-Patients Department
Western General Hospital
EH4 2XU
Tel: 0131 537 2436*

Information for patients



Review

Following fitting of your AFO, your Orthotist may arrange an appointment for review. Minor adjustments can be made to your AFO and we would recommend that your brace is reviewed by an Orthotist at least once every 12 months to ensure the device is still meeting your needs and to check for any wear and tear.

Children may need their AFO changed more often as they grow or gain weight. We would recommend that their brace is reviewed by an Orthotist at least every six months.

Never attempt to repair or adjust your brace yourself. If you experience any problems or have questions about your brace, please contact your Orthotist.

If you require a appointment with your orthotic department, and have been seen in the past 2 years, you can contact them direct without a GP referral.

Remember

Please remember to bring your AFO, when attending appointments with your Orthotist even if it is broken or no longer fits.

For a copy of this booklet in larger print, Braille, or your community language (Arabic, Bengali, Chinese, Urdu or Polish), please call 0131 536 9434.

General information

The term AFO stands for Ankle Foot Orthosis and describes the part of the body that the brace surrounds. An AFO controls the range of motion in your foot and ankle and helps to stabilize its position. Good foot and ankle alignment is vital for balancing and distributing weight when you are standing and walking.

The AFO will be designed to meet your specific needs. Each person is unique, so responses to treatment will vary. As you adjust to this new and proper alignment, you may experience some discomfort. The adjustment period usually ranges from one to six weeks when the recommended wearing guidelines are followed.

Wearing

Your AFO has been prescribed for you and is intended for your use only. It is designed to conform to the shape of your foot and ankle and may feel slightly 'strange' due to the realignment of your lower limb and pressure redistribution. It will therefore require a gentle breaking-in period.

Wear time

When you are first fitted with your AFO you should wear into it gradually. This allows you to get used to wearing the AFO and allows the skin to gradually build up resistance to the pressures which the brace is applying. You should check your skin regularly during the first few weeks. You may experience some general muscle aches until you are used to wearing the AFO.

- Start slowly, only wearing your AFO for a maximum of one hour on the first day, with low to moderate activities
- Increase the wearing time each day, by 30 minutes to one hour, until you are comfortable wearing your AFO all day

- You should increase your activity level during this transition period
- If you experience any pain or discomfort, remove the AFO from your shoes for at least two to four hours, and then try again. If pain persists, stop wearing the AFO and contact your Orthotist
- You will achieve the greatest long-term benefits when you consistently wear your AFO every day

There may be cases where your Orthotist recommends a different wearing regime, for example at night only. These instructions will be given to you when your AFO is fitted.

Shoes

- Select good shoes with proper support and always try them on with the AFO in place to ensure proper fit and comfort
- The best type of shoes to wear with your AFO feature a closed back and toe; come up over the top of the foot like laced-up sport shoes; flat shoes; a removable liner; and a slightly wider width to accommodate the AFO
- Be sure to wear stockings or long cotton socks (which are a little higher than the top of the AFO) to minimise the possibility of skin irritation
- During use, ensure that footwear is secured as tightly as is comfortable.

Skin care

Your skin should be inspected regularly, especially in the first few weeks after you have been fitted with a new AFO. A close fitting orthosis can sometimes cause some redness on the skin around the edges of the orthosis and where the AFO is applying some pressure. Check your skin thoroughly, especially over bony areas and around the edges of the AFO. If you see a red area that does not go away within 30 minutes of removing the AFO, or notice a break in the skin or blisters, you should stop wearing the AFO and make an appointment with your Orthotist to have it adjusted.

Caring for your AFO

Cleaning

Use an antibacterial soap and warm water to wipe the inside surfaces of your AFO. Towel dry the AFO completely before putting it back on. Avoid drying in direct sources of heat (such as on the heater or using a hair dryer) as this can deform the plastic.

Straps and Padding

The straps and padding may become worn and can be replaced. To arrange repairs please contact the Orthotic department.

Plastic

Inspect the plastic regularly, checking for any cracks or areas of fatigue. Never attempt to repair, trim or adjust the AFO yourself. If a problem arises, contact the Orthotics department immediately.

Joints

Some AFOs may have joints at the ankle. These should be checked before each time the AFO is used looking for any dents or cracks and to make sure that all rivets and screws are secure.